

Certificate of Analysis

Distilled Water

All tests below are carried out at 20°C and
is clear, colourless and odourless while boiling

| IMPURITY | RESULTS |
|------------------------|-----------|
| Residue on evaporation | <1.0mg/L |
| Ammonium Nitrogen | <0.1mg/L |
| Nitrate Nitrogen | <0.1mg/L |
| Nitrite Nitrogen | <0.1mg/L |
| Chloride | <1.0mg/L |
| Copper | <0.01mg/L |
| Iron | <0.01mg/L |
| Manganese | <0.01mg/L |
| Aluminium | <0.1mg/L |
| Chlorines | <0.1mg/L |
| Sulfates | <0.1mg/L |
| Conductivity | 0.63 ms/m |
| Permanganate Mn-COD | <1mg/L |
| pH | 6.3 |

*The tests contained in this report have been carried out on the samples as
received by Watertest, in accordance with APHA Standard Methods
of Water and Wastewater 23rd Edition*